



No: PUR/2009-10/IND10783/RSP

Date: 01/12/2009

To

Sub : Request for Pro-forma Invoice

Dear Sir

This Research Institute is interested in purchasing of the items mentioned below. You are therefore requested to forward a pro-forma invoice along with your terms & conditions of supply & payment. Please provide a copy of the technical literature and the specifications of these items in a sealed envelope.

Please read carefully the terms and conditions mentioned overleaf and submit your pro-forma invoice/quotation accordingly.

SNo.	Name of Item	Catalog No	Qty
1	Rotational Vacuum concentrator, 1		1

Important

- Please mention our reference number on the envelop and address all correspondence to Director IMT Chandigarh only.
- Please quote strictly as per our specifications.
- Please attach compliance sheet
- Proforma Invoice received after due date will not be considered under any circumstances.
- Please note that your Quotation should reach us latest by **15/12/2009** opened on **16/12/2009**
- **Your Quotation must be duly stamped and reach directly to IMT from you only. If quotation is submitted by Indian representative/agent then they must have to produce a authority certificate of principal party for quotating the price Other wise it will be very difficult at our end to consider your quotation.**

Yours faithfully,

Stores & Purchase Officer.

Specifications for a Rotational Vacuum concentrator: Sep 25, 2009

1. Teflon-coated chamber with digital (LED / LCD) display
2. (A) Teflon coated chamber, and (B) Acrylic/glass lid with safety interlock
3. Instrument temperature 35-65°C during use with high temperature auto cut-off feature
4. Lowest operating time (timer adjustment) should be at least 2 minutes or lower and highest time should be at least five hours or higher
5. Rotor for at least (A) 24 x 1.5/2.0 ml tubes and (B) speed range from 1200-1400 rpm
6. The concentrator should have (A) universal vacuum source of 30-40 L displacement, and (B) vacuum strength of 8-13 mbar
7. The concentrator should have (A) oil-free maintenance, (B) maintenance-free Teflon diaphragm, and (C) chemical compatibility with aqueous and alcohol solvents

Optional conditions

The supplier must provide the following items along with quotation

1. Compliance statement (with literature) in the tabular form as per tender specifications
2. Warranty of at least one year and AMC charge after one year
3. List of users
4. List and price of minimally recommended spares for 5 years